



RAN - 1903001103020001

RAN-1903001103020001



B. Sc. (Biotechnology) (Sem. - III) Examination

March - 2023

BT-05 : Instrumentations and Techniques

સૂચના : / Instructions

(1)

નીચે દર્શાવેલ નિશાનીવાળી વિગતો ઉત્તરવહી પર અવશ્ય લખવી.
Fill up strictly the details of signs on your answer book

Name of the Examination:

B. Sc. (Biotechnology) (Sem. - III)

Name of the Subject :

BT-05 : Instrumentations and Techniques

Subject Code No.: **1903001103020001**

Seat No.:

<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
----------------------	----------------------	----------------------	----------------------	----------------------	----------------------

Student's Signature

(2) All questions are compulsory.

***O.M.R. Sheet ભરવા અંગેની અગત્યની સૂચનાઓ આપેલ
O.M.R. Sheetની પાછળ છાપેલ છે.***

***Important instructions to fillup O.M.R. Sheet
are given on back side of the provided O.M.R. Sheet.***

- Q. 1. Alkalinity of a solution _____ as the pH _____.
- A. Decreases, increases B. Increases, increases
C. Increases, decreases D. None of the above
- Q. 2. Which one is not included in planer chromatography?
- A. TLC B. Paper Chromatography
C. HPLC D. HPTLC
- Q. 3. One electron volt (eV) is equal to _____.
- A. 1.6×10^{-16} J B. 1.9×10^{-19} J
C. 1.6×10^{-19} J D. 1.9×10^{-16} J
- Q. 4. Cocktails used in scintillation fluids does not contain _____.
- A. 2,5-diphenyloxazole
B. 1,4 bis(5phenyloxazol-2-yl) benzene
C. 2-(4'-t-butylphenyl)-5-(4''-biphenyl)-1,3,4,-oxydiazole
D. 1,4 bis(5benzenoxazol-2-yl) phenyl
- Q. 5. Intensifying screen sheet is used in _____.
- A. Direct autoradiography
B. Indirect autoradiography
C. Converged autoradiography
D. Conservational autoradiography
- Q. 6 _____ is defined as the total amount, as distinct from the concentration, of analyte present in one phase divided by the total amount present in the other phase.
- A. Distribution coefficient
B. Partition coefficient
C. Effective distribution coefficient
D. Concentration

- Q. 7.** Choose the correct statement for chromatography.
- A. Higher the adsorption to the stationary phase, the faster the molecule will move through the column.
 - B. Higher the solubility in the mobile phase, the faster the molecule will move through the column
 - C. Higher the absorption to the stationary phase, the faster the molecule will move through the column
 - D. None of the above are correct.
- Q. 8.** Pre-activation of chromatographic plate can be carried out at which temperature in Hot air oven?
- A. 101 °C
 - B. 301 °C
 - C. 401 °C
 - D. 201 °C
- Q. 9.** Which of the following is an effective way of purifying liquids containing suspensions?
- A. Crystallization
 - B. Decanting
 - C. Centrifuging
 - D. Separating funnel
- Q. 10.** Which type of rotors are useful for gradient centrifugation of biological elements that do not properly participate in conventional gradient?
- A. Fixed angle
 - B. Vertical
 - C. Swinging bucket
 - D. Near vertical
- Q. 11.** Which of the following can be used to establish gradient for the separation of whole cells and sub-cellular particles?
- A. Percoll
 - B. Ficoll
 - C. Dextran
 - D. All
- Q. 12.** Give the mathematical equation to describe Stock's law.
- A. $v = 2/9 r^2 (\rho_p - \rho_m) / \eta \times g$
 - B. $v = 9/2 r^2 (\rho_p - \rho_m) / \eta \times g$
 - C. $v = 2/9 \eta \times g / r^2 (\rho_p - \rho_m)$
 - D. None

- Q. 20.** After centrifugation, sublimate _____.
- A. Dissolves completely
 - B. Settles at bottom
 - C. Depends upon pH of sublimate
 - D. Remain suspended in a liquid
- Q. 21.** A grating consists of ruled lines, as many as _____ line per mm, on a transparent or reflecting base.
- A. 20
 - B. 200
 - C. 500
 - D. 2000
- Q. 22.** Which of the below lamp can be used as light source for U.V. light?
- A. The Deuterium Lamp
 - B. Hydrogen Discharge Lamp
 - C. Tungsten Halogen Lamp
 - D. All of the given
- Q. 23.** Which of the given option correctly represent the dimensions of the cuvette?
- A. 1.00 cm by 1.00 cm cross section and are several centimeters in height
 - B. 0.10 cm by 1.00 cm cross section and are 1.00 centimeters in height
 - C. 10.0 cm by 10.0 cm cross section and are 1.0 centimeters in height
 - D. 1.00 cm by 10.0 cm cross section and are several centimeters in height
- Q. 24.** Function of reference beam is to _____.
- A. Provide the final light intensity after passing through the sample
 - B. Provide the initial light intensity before passing through the sample
 - C. Both of the given
 - D. None of the given
- Q. 25.** The colorimeter is generally allowed to warm up for 15 minutes in order to _____.
- A. Perform auto-cleaning
 - B. Generate Vacuum
 - C. Stabilize the light source and the detector
 - D. None of the given

- Q. 26.** The useful range of absorbance scale is from 0-2 but it is desirable to keep within the range 0-1 because, above 1, the results become unreliable due _____.
- A. To scattering of light.
 - B. To reflection of light from the edge of solution
 - C. To difficulty in calculation results
 - D. All of the given
- Q. 27.** In colorimeter, process of recording “zero absorbance” is called as _____.
- A. Celebration
 - B. Calibration
 - C. Zeroing
 - D. Auto-tuning
- Q. 28.** Beer-Lambert’s law is only applicable to _____.
- A. A beam of monochromatic light passes through a homogeneous transparent medium
 - B. A beam of dichromatic light passes through a homogeneous transparent medium
 - C. A beam of multichromatic light passes through a homogeneous transparent medium
 - D. All of the given are correct
- Q. 29.** In U.V./VIS spectrometry, _____ acts as the silent component of the sample.
- A. Water
 - B. Proteins
 - C. Nucleic Acids
 - D. Organic Solvents
- Q. 30.** The absorption of UV light by nucleic acids depends upon _____.
- A. $n-\pi^*$ transition of purine
 - B. $\pi-\pi^*$ transition of purine and pyrimidine
 - C. $n-\pi^*$ and $\pi-\pi^*$ transition of purine and pyrimidine
 - D. $n-n^*$, $n-\pi^*$ and $\pi-\pi^*$ transition of purine and pyrimidine
- Q. 31.** Molecular sub-structures responsible for interaction with electromagnetic radiation are called _____.
- A. Chromophores
 - B. Fluorophores
 - C. Chromogens
 - D. Fluorogens

- Q. 38.** Lanthanum fluoride is used in _____ electrode.
- A. Lanthanum selective electrode
 - B. Fluoride selective electrode
 - C. Enzyme electrode
 - D. Calcium selective electrode
- Q. 39.** pH meter is a _____ which measures the voltage between two electrodes placed in a solution.
- A. Potentiometer
 - B. Voltmeter
 - C. Conductometer
 - D. Thermometer
- Q. 40.** In liquid membrane electrode, the liquid ion exchanger is held in a porous disc of _____.
- A. Solid material
 - B. Semi-permeable membrane
 - C. Hydrophobic material
 - D. Water absorbing material
- Q. 41.** A slice of rabbit liver was used to construct _____ electrode?
- A. Guanine selective
 - B. Calcium selective
 - C. Fluoride selective
 - D. None of given
- Q. 42.** Why does hydroxide ion interfere with the measurement of fluoride ion?
- A. Large size
 - B. Small size
 - C. Same size
 - D. Unknown reason
- Q. 43.** Which solution is filled in calomel electrode?
- A. Saturated solution of potassium chloride
 - B. Saturated solution of potassium iodide
 - C. Saturated solution Sodium chloride
 - D. Saturated solution of aluminium chloride
- Q. 44.** Which of the following is not the characteristic of ion selective electrodes?
- A. It is fragile
 - B. Easy to use
 - C. Available in different sizes and shapes
 - D. It is insensitive to many ions

- Q. 45.** In recent liquid membrane electrodes, the porous liquid membrane is replaced with which of the following?
- A. Polyvinyl chloride B. Polyacryl chloride
C. Polyester membrane D. Polyacryl amide
- Q. 46.** Identify the radioactive version of carbon.
- A. ^{12}C B. ^{15}C
C. ^{13}C D. ^{14}C
- Q. 47.** What does act as an anode in Geiger Muller counter?
- A. Argon gas B. Platinum rod
C. Copper cylinder D. Steel plate
- Q. 48.** Chromatography technique is invented by _____.
- A. M. Tswett B. CremerC.
C. William Harris D. Duboscq
- Q. 49.** Which of the Factors affects the efficiency of autoradiography?
- A. Thickness of the sample B. All of Given
C. Thickness of emulsion D. Exposure time
- Q. 50.** Identify SI unit of radioactivity.
- A. Curie B. Rutherford
C. Becquerel D. Milicurie
-

SPACE FOR ROUGH WORK